

Biphasic Plate Distal Femur (BPDF)

Fast, robust healing of long-bone fractures through confident weight bearing is the goal of the Biphasic Plating concept. In collaboration with the AO's Innovation Funding and the AO Research Institute, Davos in Switzerland, and Queensland University of Technology (QUT) in Brisbane, Australia.

Intended Use

The Biphasic Plate DF is intended for temporary fixation, correction or stabilization of the distal femur.

Indications

Stabilization of fractures of the distal femur, this includes:

- Distal shaft fractures
- Supracondylar fractures
- Intra-articular condylar fractures
- Periprosthetic fractures

Contraindications

- Pregnant woman
- Nursing woman
- Systemic infection
- Hypersensitivity to material

Potential Side Effects and Risks please see instruction for use (IFU) on our website.

Compatibility

- DePuy Synthes Locking Screw \varnothing 5.0 mm
- DePuy Synthes Locking Screw \varnothing 3.5 mm
- DePuy Synthes Instruments

Materials

Stainless steel 1.4441 ASTM F138

Articles

- 100.178 Biphasic Plate DF Left
- 100.179 Biphasic Plate DF Right

In cooperation with:



41medical

Tannlimattweg 14 · 2544 Bettlach · Switzerland

T +41 32 645 41 41 · F +41 32 645 09 59 · info@41medical.com · www.41medical.com

Scan for more
information
[41medical.com](https://www.41medical.com)

